

SUSPECT ADVERSE REACTION REPORT

I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) PRIVACY	1a. COUNTRY DOMINICAN REPUBLIC	2. DATE OF BIRTH			2a. AGE 11 Years	3. SEX Male	3a. WEIGHT Unk	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION <input type="checkbox"/> PATIENT DIED <input type="checkbox"/> INVOLVED OR PROLONGED INPATIENT HOSPITALISATION <input type="checkbox"/> INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY <input type="checkbox"/> LIFE THREATENING
		Day	Month	Year				Day	Month	Year	
			PRIVACY					Unk			

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)
when the dose was set the device administered more than the quantity set [Incorrect dose administered by device]
when the dose was set the device administered more than the quantity set [Device delivery system issue]
medication continued to come out and it came out of the patient's skin. [Accidental exposure to product related to device use]
medication continued to come out and it came out of the patient's skin.

(Continued on Additional Information Page)

II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LR7824; Exp.Dt. JUN-2027} #2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection {Lot # LD7551}		20. DID REACTION ABATE AFTER STOPPING DRUG? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
15. DAILY DOSE(S) #1) 1 mg, 1x/day (at night) #2)	16. ROUTE(S) OF ADMINISTRATION #1) Unknown #2) Unknown	
17. INDICATION(S) FOR USE #1) Unknown #2) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
18. THERAPY DATES(from/to) #1) Unknown #2) Unknown	19. THERAPY DURATION #1) Unknown #2) Unknown	

III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates Type of History / Notes Description Unknown		

IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San Jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. PV202500057328	
24c. DATE RECEIVED BY MANUFACTURER 09-JUL-2025	24d. REPORT SOURCE <input type="checkbox"/> STUDY <input type="checkbox"/> LITERATURE <input checked="" type="checkbox"/> HEALTH PROFESSIONAL <input checked="" type="checkbox"/> OTHER: Spontaneous	
DATE OF THIS REPORT 14-JUL-2025	25a. REPORT TYPE <input type="checkbox"/> INITIAL <input checked="" type="checkbox"/> FOLLOWUP: 3	

25b. NAME AND ADDRESS OF REPORTER
NAME AND ADDRESS WITHHELD.

NAME AND ADDRESS WITHHELD.

NAME AND ADDRESS WITHHELD.

ADDITIONAL INFORMATION**7+13. DESCRIBE REACTION(S) continued**

[Exposure via skin contact]

the case was leaking the medication [Device leakage]

the brown gummy that indicates the doses administered did not go down and continued to stay with the air [Device failure to prime]

Case Description: This is a spontaneous report received from a Nurse and a Consumer or other non HCP, Program ID: 164974.

An 11-year-old male patient received somatropin (GENOTROPIN PEN), (Lot number: LR7824, Expiration Date: Jun2027) at 1 mg 1x/day (1 mg, 1x/day (at night)), Device Lot Number: LD7551, Device Expiration Date: Jan2027. The patient's relevant medical history and concomitant medications were not reported.

The following information was reported: INCORRECT DOSE ADMINISTERED BY DEVICE (non-serious), DEVICE DELIVERY SYSTEM ISSUE (non-serious) and all described as "when the dose was set the device administered more than the quantity set"; ACCIDENTAL EXPOSURE TO PRODUCT (non-serious), EXPOSURE VIA SKIN CONTACT (non-serious) and all described as "medication continued to come out and it came out of the patient's skin."; DEVICE LEAKAGE (non-serious), described as "the case was leaking the medication"; DEVICE FAILURE (non-serious), described as "the brown gummy that indicates the doses administered did not go down and continued to stay with the air". The action taken for somatropin was unknown.

Causality for "when the dose was set the device administered more than the quantity set", "medication continued to come out and it came out of the patient's skin.", "the case was leaking the medication" and "the brown gummy that indicates the doses administered did not go down and continued to stay with the air" was determined associated to device constituent of somatropin (malfunction).

Additional information: Reporter stated that when the dose was set the device administered more than the quantity set, it did not stay at the milligrams set, so when 1 mg was set, it did not administer that. On 23May2025 nurse reported that caregiver did not take the pen to the appointment so it was not possible to correct the turns she gives. Caregiver believed it could be applying maybe more, maybe less. As of 04Jul2025, patient's caregiver reported that when administering the medication to the patient, she inserted the vial into the case, but the case was leaking the medication, she did all the steps and the respective procedure, she eliminated the air, but she saw that it still had air. Last night, when she gave the dose to the patient and it was leaking medicine through the needle, and when she inserted it into the child, the brown gummy that indicates the doses administered did not go down and continued to stay with the air. So instead of inserting everything with the needle in place (count the 10 seconds after applying it) the medication continued to come out and it came out of the patient's skin. Reporter was given a the contact of a nurse and was advised not to administer the medication until it was checked by the nurse. Upon a follow-up received on 09Jul2025, nurse stated that at first, there was an issue with the pen, but upon closer inspection, she realized the problem was due to the medication, which had a very small amount left, making it difficult for it to dispense. She tried several times and eventually identified the cause. In the end, the patient was satisfied.

Follow-up (23May2025). This is a spontaneous follow-up report received from a nurse. Updated information included: clinical course.

Follow-up (27Jun2025): Follow-up attempts are completed. Batch/lot number is not provided, and it cannot be obtained.

Follow-up (04Jul2025). This is a spontaneous follow-up report received from Consumer or other non HCP, Program ID: 164974.

Updated information includes: new events of "accidental exposure to product", "exposure via skin contact", device leakage and device failure to prime, additional information.

Follow-up (09Jul2025). This is a spontaneous follow-up report received from a nurse, Program ID: 164974.

Updated information included: clinical course.