

SUSPECT ADVERSE REACTION REPORT

I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) PRIVACY	1a. COUNTRY DOMINICAN REPUBLIC	2. DATE OF BIRTH			2a. AGE 2 Years	3. SEX Female	3a. WEIGHT Unk	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION <input type="checkbox"/> PATIENT DIED <input type="checkbox"/> INVOLVED OR PROLONGED INPATIENT HOSPITALISATION <input type="checkbox"/> INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY <input type="checkbox"/> LIFE THREATENING
		Day	Month	Year				Day	Month	Year	
			PRIVACY					02	JUL	2025	

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)
changed the cartridge as if it came out a lot (leakage) [Device leakage]
it looked like lots of bubbles inside [Device physical property issue]

Case Description: This is a spontaneous report received from a Consumer or other non HCP from product quality group, Program ID: 164974..

A 2-year-old female patient received somatropin (GENOTROPIN PEN), (Lot number: LR7824, Expiration Date: Jun2027) at 0.3 mg 1x/day (0.3 mg, 1x/day (at night)), Device Expiration Date: Jan2027. The patient's relevant medical history and concomitant medications were not reported.

(Continued on Additional Information Page)

II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LR7824; Exp.Dt. JUN-2027} #2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection		20. DID REACTION ABATE AFTER STOPPING DRUG? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
15. DAILY DOSE(S) #1) 0.3 mg, 1x/day (at night) #2)	16. ROUTE(S) OF ADMINISTRATION #1) Unknown #2) Unknown	
17. INDICATION(S) FOR USE #1) Unknown #2) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
18. THERAPY DATES(from/to) #1) Unknown #2) Unknown	19. THERAPY DURATION #1) Unknown #2) Unknown	

III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates Type of History / Notes Description Unknown		

IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San Jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. PV202500081152	25b. NAME AND ADDRESS OF REPORTER NAME AND ADDRESS WITHHELD. NAME AND ADDRESS WITHHELD.
24c. DATE RECEIVED BY MANUFACTURER 21-AUG-2025	24d. REPORT SOURCE <input type="checkbox"/> STUDY <input type="checkbox"/> LITERATURE <input type="checkbox"/> HEALTH PROFESSIONAL <input checked="" type="checkbox"/> OTHER: Spontaneous	
DATE OF THIS REPORT 27-AUG-2025	25a. REPORT TYPE <input type="checkbox"/> INITIAL <input checked="" type="checkbox"/> FOLLOWUP: 1	

27-Aug-2025 10:54

ADDITIONAL INFORMATION**7+13. DESCRIBE REACTION(S) continued**

The following information was reported: DEVICE LEAKAGE (non-serious) with onset 02Jul2025, outcome "unknown", described as "changed the cartridge as if it came out a lot (leakage)"; DEVICE PHYSICAL PROPERTY ISSUE (non-serious) with onset 02Jul2025, outcome "unknown", described as "it looked like lots of bubbles inside".

Causality for "changed the cartridge as if it came out a lot (leakage)" and "it looked like lots of bubbles inside" was determined associated to device constituent of somatropin (malfunction).

Product Quality Group provided investigational summary and conclusion on 21Aug2025 for somatropin (device constituent): Site investigation (Puurs): Container Leaking During Prep/Use The complaint for "changed the cartridge as if it came out a lot" of Genotropin Pen Injectable was investigated. The investigation included reviewing an analysis of the complaint history for the involved scope and Annual Product Review. A complaint sample was not returned. The complaint is not confirmed. No root cause or CAPA were identified as the complaint was not confirmed. No related quality issues were identified during the investigation. There is no impact on product quality, regulatory, validation, stability and patient safety. The Issue Escalation (NTM) process determined that no regulatory notification was required. The final scope was determined to be the reported product and product type, as no lot was available. Device Investigation: This investigation is based on the information captured in the Complaint Description and Argus Report. Two distinct Complaint Issues of "Cannot Remove Air" and "Leaking During Loading" were reported. However, these two distinct Complaint Issues map to same Hazard/Hazardous Situation. The Risk Management File was reviewed to confirm that the Hazard(s) and Hazardous Situation(s) associated with the Complaint Issue are documented in the Hazard Analysis (INX# 100281795, Version # (9.0)). All complaint investigations are trended. There is no current trend alert documented.

Product Quality Group provided investigational summary and conclusion on 22Aug2025 for somatropin (device constituent): MDCP Investigation Summary and Conclusion: The complaint of The patient's mother says, "My daughter received the medication genotropin and on Wednesday, July 2, I made the change for the first time, that is, after advice, I changed the cartridge as if it came out a lot (leakage) and it stayed on the medication and it looked like lots of blisters inside, and when I tried to remove the air with the pen it didn't come out, so now the medication got blistered, I don't know if it's normal or should do it." Proceed with nurse advisory coordination for Genotropin Pen 5.3 was investigated by the manufacturing site. Site investigation (Puurs): Container Leaking During Prep/Use The complaint for "changed the cartridge as if it came out a lot" of Genotropin Pen Injectable was investigated. The investigation included reviewing an analysis of the complaint history for the involved scope and Annual Product Review. A complaint sample was not returned. The complaint is not confirmed. No root cause or CAPA were identified as the complaint was not confirmed. No related quality issues were identified during the investigation. There is no impact on product quality, regulatory, validation, stability and patient safety. The Issue Escalation (NTM) process determined that no regulatory notification was required. The final scope was determined to be the reported product and product type, as no lot was available. Device Investigation: This investigation is based on the information captured in the Complaint Description and Argus Report. Two distinct Complaint Issues of "Cannot Remove Air" and "Leaking During Loading" were reported. However, these two distinct Complaint Issues map to same Hazard/Hazardous Situation. The Risk Management File was reviewed to confirm that the Hazard(s) and Hazardous Situation(s) associated with the Complaint Issue are documented in the Hazard Analysis (INX# 100281795, Version # (9.0)). All complaint investigations are trended. There is no current trend alert documented.

Additional information: The patient's mother stated "My daughter received the medication Genotropin and on Wednesday, 02Jul2025, they changed the cartridge as if it came out a lot (leakage) and it stayed on the medication and it looked like lots of bubbles inside, and when they tried to remove the air with the pen it didn't come out, so now the medication got bubbles. Pen lot number: PAA141175 Follow-up (22Aug2025): Follow-up attempts are completed.

Follow-up(21Aug2025 and 22Aug2025): This is a follow-up report from product quality group providing investigation results. Updated information: Investigation results updated.