

# SUSPECT ADVERSE REACTION REPORT

## I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) <b>PRIVACY</b>	1a. COUNTRY <b>GUATEMALA</b>	2. DATE OF BIRTH			2a. AGE <b>15 Years</b>	3. SEX <b>Female</b>	3a. WEIGHT <b>Unk</b>	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION
		Day	Month	Year				Day	Month	Year	
		<b>PRIVACY</b>						<b>Unk</b>			

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)  
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)  
the trigger button doesn't work well [Device mechanical jam]  
device was not working [Device defective]  
Yesterday we couldn't give the girl the medication [Drug dose omission by device]

Case Description: This is a spontaneous report and received from Consumer or other non HCPs from product quality group, Program ID: 164974.

A 15-year-old female patient received somatropin (GENOTROPIN PEN),

(Continued on Additional Information Page)

☐ PATIENT DIED  
☐ INVOLVED OR PROLONGED INPATIENT HOSPITALISATION  
☐ INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY  
☐ LIFE THREATENING

## II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1 ) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LR7825; Exp.Dt. MAY-2027} #2 ) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection		20. DID REACTION ABATE AFTER STOPPING DRUG?  <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
15. DAILY DOSE(S) #1 ) 2 mg, daily (every night) #2 )	16. ROUTE(S) OF ADMINISTRATION #1 ) Unknown #2 ) Unknown	
17. INDICATION(S) FOR USE #1 ) Unknown #2 ) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION?  <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
18. THERAPY DATES(from/to) #1 ) Unknown #2 ) Unknown	19. THERAPY DURATION #1 ) Unknown #2 ) Unknown	

## III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates                      Type of History / Notes                      Description Unknown		

## IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. <b>PV202400136352</b>	
24c. DATE RECEIVED BY MANUFACTURER <b>16-MAY-2025</b>	24d. REPORT SOURCE <input type="checkbox"/> STUDY <input type="checkbox"/> LITERATURE <input type="checkbox"/> HEALTH PROFESSIONAL <input checked="" type="checkbox"/> OTHER: Spontaneous	
DATE OF THIS REPORT <b>21-MAY-2025</b>	25a. REPORT TYPE <input type="checkbox"/> INITIAL <input checked="" type="checkbox"/> FOLLOWUP: 6	
		25b. NAME AND ADDRESS OF REPORTER NAME AND ADDRESS WITHHELD.  NAME AND ADDRESS WITHHELD.  NAME AND ADDRESS WITHHELD.

21-May-2025 16:00

---

**ADDITIONAL INFORMATION****7+13. DESCRIBE REACTION(S) continued**

(Lot number: LR7825, Expiration Date: May2027) at 2 mg daily (2 mg, daily (every night)). The patient's relevant medical history and concomitant medications were not reported.

The following information was reported: DEVICE MECHANICAL ISSUE (non-serious), described as "the trigger button doesn't work well"; DEVICE DEFECTIVE (non-serious), described as "device was not working"; DRUG DOSE OMISSION BY DEVICE (non-serious), described as "Yesterday we couldn't give the girl the medication". The action taken for somatropin was unknown.

Causality for "the trigger button doesn't work well", "device was not working" and "yesterday we couldn't give the girl the medication" was determined associated to device constituent of somatropin (malfunction).

Product Quality Group provided investigational results on 28Nov2024, 02Mar2025 and 16May2025 for somatropin (device constituent): Investigation Summary and Conclusion: No further investigation was required as no valid lot number or returned sample was available. This complaint will continue to be trended. If additional information becomes available, this complaint will be reopened. Device Investigation: This investigation is based on the information captured in the Complaint Description and Argus Report. The Complaint Issue, Injection Knob/Dial Issue and Loss of Function were reported. The Risk Management File was reviewed to confirm that the Hazard(s) and Hazardous Situation(s) associated with the Complaint Issue are documented in the Hazard Analysis (INX100281795), Version (9.0). All complaint investigations are trended. There is not a current trend alert documented.

Additional Information: On 28Apr2025, the patient's caregiver indicated that the device was not working and mentioned that they were on a cruise. A virtual consultation was offered, but the patient's mother requested a prescription to give to the ship's staff to see if they had any needles.

Follow-up (28Nov2024): This is a follow-up report from product quality group providing investigation results.

Follow-up (02Mar2025): This is a follow-up report from product quality group providing investigation results.

Follow-up (28Apr2025): This is a follow-up report from received from a Consumer or other non HCP, Program ID: 164974. Updated information: new reporters, new event added ("device was not working"), product details (dosage regimen, drug lot number and drug expiration date) and clinical course.

Follow-up (07May2025): This is a follow-up report from product quality group. Updated information: device details (medical device component code).

Follow-up (16May2025): This is a spontaneous follow-up report from product quality group. Updated information: investigation results