

SUSPECT ADVERSE REACTION REPORT

I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) PRIVACY	1a. COUNTRY GUATEMALA	2. DATE OF BIRTH			2a. AGE 3 Years	3. SEX Female	3a. WEIGHT Unk	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION ☐ PATIENT DIED ☐ INVOLVED OR PROLONGED INPATIENT HOSPITALISATION ☐ INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY ☐ LIFE THREATENING
		Day	Month	Year				Day	Month	Year	
			PRIVACY					14	MAY	2025	

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)
it's been three days with no medication coming out. [Drug dose omission by device]
the medication did not come out of the needle [Device blockage]
that the plunger of the pen did not return to the top, that is why a certain part of the cartridge spilled, that is, She did not make the replacement correctly [Device handling error]

Case Description: This is a spontaneous report received from a Consumer or other non HCP, Program ID: 164974.

(Continued on Additional Information Page)

II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LN4285; Exp.Dt. FEB-2027} #2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution (Continued on Additional Information Page)		20. DID REACTION ABATE AFTER STOPPING DRUG? ☐ YES ☐ NO ☒ NA
15. DAILY DOSE(S) #1) 0.6 mg, daily (at night) #2)	16. ROUTE(S) OF ADMINISTRATION #1) Unknown #2) Unknown	
17. INDICATION(S) FOR USE #1) Unknown #2) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION? ☐ YES ☐ NO ☒ NA
18. THERAPY DATES(from/to) #1) Unknown #2) Unknown	19. THERAPY DURATION #1) Unknown #2) Unknown	

III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates Type of History / Notes Description Unknown		

IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. PV202500058000	25b. NAME AND ADDRESS OF REPORTER NAME AND ADDRESS WITHHELD.
24c. DATE RECEIVED BY MANUFACTURER 15-MAY-2025	24d. REPORT SOURCE ☐ STUDY ☐ LITERATURE ☐ HEALTH PROFESSIONAL ☒ OTHER: Spontaneous	NAME AND ADDRESS WITHHELD.
DATE OF THIS REPORT 21-MAY-2025	25a. REPORT TYPE ☒ INITIAL ☐ FOLLOWUP:	

21-May-2025 16:23

ADDITIONAL INFORMATION**7+13. DESCRIBE REACTION(S) continued**

A 3-year-old female patient received somatropin (GENOTROPIN PEN), first regimen (Lot number: LN4285, Expiration Date: Feb2027) at 0.6 mg daily (0.6 mg, daily (at night)) and second regimen (Lot number: LD7549, Expiration Date: Mar2027) at 0.6 mg daily (0.6 mg, daily (at night)), Device Lot Number: D154, Device Expiration Date: Mar2027. The patient's relevant medical history and concomitant medications were not reported.

The following information was reported: DEVICE USE ERROR (non-serious) with onset 14May2025, described as "that the plunger of the pen did not return to the top, that is why a certain part of the cartridge spilled, that is, She did not make the replacement correctly"; DRUG DOSE OMISSION BY DEVICE (non-serious), described as "it's been three days with no medication coming out."; DEVICE OCCLUSION (non-serious), described as "the medication did not come out of the needle". The action taken for somatropin was unknown.

Causality for "it's been three days with no medication coming out.", "the medication did not come out of the needle" and "that the plunger of the pen did not return to the top, that is why a certain part of the cartridge spilled, that is, she did not make the replacement correctly" was determined associated to device constituent of somatropin (malfunction).

Additional information: On 15May2025: Nurse indicates: "In face-to-face counseling with this patient who went at two in the afternoon of May 14, the person in charge reports that she made a cartridge change and a certain part of the cartridge spilled in the refill. But it is observed that the plunger of the pen did not return to the top, that is why a certain part of the cartridge spilled, that is, She did not make the replacement correctly. She forgot the step of returning the plunger of the pen until it reached the top counterclockwise, it is important considering that the person in charge had already made three previous replacements, so she is given advice again from the beginning and the revision of the pen is carried out with a placebo again. The placebo and PEN is in perfect condition for use. It is worth mentioning that the person in charge of poor communication from the Social Security here in Guatemala never communicated with the Pfizer Me Cuida call center for the scheduling of the advice, that is, she did it on her own, yesterday she already received advice and obviously she did tell me that she was doing the procedure wrong."

Follow-up (15May2025): This is a spontaneous follow-up report received from a Nurse, Program ID: 164974. Updated information includes: new dosage regimen, new event (Device use error) and clinical course details.

14-19. SUSPECT DRUG(S) continued

14. SUSPECT DRUG(S) (include generic name)	15. DAILY DOSE(S); 16. ROUTE(S) OF ADMIN	17. INDICATION(S) FOR USE	18. THERAPY DATES (from/to); 19. THERAPY DURATION
#1) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LD7549; Exp.Dt. MAR-2027}; Regimen #2	0.6 mg, daily (at night); Unknown	Unknown	Unknown; Unknown
#2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection {Lot # D154}; Regimen #1	; Unknown	Unknown	Unknown; Unknown