

SUSPECT ADVERSE REACTION REPORT

I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) PRIVACY	1a. COUNTRY GUATEMALA	2. DATE OF BIRTH			2a. AGE 11 Years	3. SEX Female	3a. WEIGHT Unk	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION <input type="checkbox"/> PATIENT DIED <input type="checkbox"/> INVOLVED OR PROLONGED INPATIENT HOSPITALISATION <input type="checkbox"/> INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY <input type="checkbox"/> LIFE THREATENING
		Day	Month	Year				Day	Month	Year	
		PRIVACY						Unk			

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)
the button to apply the Genotropin device is hard [Device mechanical jam]

Case Description: This is a spontaneous report received from a Consumer or other non HCP and a Nurse from product quality group, Program ID: 164974.

An 11-year-old female patient received somatropin (GENOTROPIN PEN), (Lot number: LJ9497, Expiration Date: Jan2027) at 1.4 mg 1x/day, Device Lot Number: KC8721, Device Expiration Date: 31Jan2027. The patient's relevant medical history and concomitant medications were not reported.

(Continued on Additional Information Page)

II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1) Genotropin Pen (SOMATROPIN) Solution for injection {Lot # LJ9497; Exp.Dt. JAN-2027} #2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection {Lot # KC8721}		20. DID REACTION ABATE AFTER STOPPING DRUG? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
15. DAILY DOSE(S) #1) 1.4 mg, 1x/day #2) 1.4 mg, 1x/day	16. ROUTE(S) OF ADMINISTRATION #1) Unknown #2) Unknown	
17. INDICATION(S) FOR USE #1) Unknown #2) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
18. THERAPY DATES(from/to) #1) Unknown #2) Unknown	19. THERAPY DURATION #1) Unknown #2) Unknown	

III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates Type of History / Notes Description Unknown		

IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. PV202500085583	
24c. DATE RECEIVED BY MANUFACTURER 24-JUL-2025	24d. REPORT SOURCE <input type="checkbox"/> STUDY <input type="checkbox"/> LITERATURE <input type="checkbox"/> HEALTH PROFESSIONAL <input checked="" type="checkbox"/> OTHER: Spontaneous	
DATE OF THIS REPORT 04-AUG-2025	25a. REPORT TYPE <input checked="" type="checkbox"/> INITIAL <input type="checkbox"/> FOLLOWUP:	

25b. NAME AND ADDRESS OF REPORTER
NAME AND ADDRESS WITHHELD.

 NAME AND ADDRESS WITHHELD.

 NAME AND ADDRESS WITHHELD.

ADDITIONAL INFORMATION**7+13. DESCRIBE REACTION(S) continued**

The following information was reported: DEVICE MECHANICAL ISSUE (non-serious), outcome "unknown", described as "the button to apply the Genotropin device is hard".

Additional Information: Patient caregiver informs you that they want the nurse visit as soon as possible since the button to apply the Genotropin device is hard. Later, the patient asked for another consultation to check the pen again, and it was confirmed that the issue was not with the pen, but with the cartridges, as they were purchased in Panama and brought to Guatemala. The problem with the pen was that the cartridges were not in perfect condition (they had abnormalities) and did not fit properly into the device. Due that, the pen was malfunctioning. The nurse advised the family not to use the medication anymore.

Product Quality Group provided investigational results on 28Jul2025 for somatropin (device constituent): Investigation Summary and Conclusion: Site investigation (Pfizer Manufacturing Site): No further investigation was required as no valid lot number or returned sample was available. This complaint will continue to be trended. If additional information becomes available, this complaint will be reopened. Device engineering investigation: This investigation is based on the information captured in the Complaint Description and Argus Report. The Complaint Issue, Injection Knob/Dial Issue, was reported. The Risk Management File was reviewed to confirm that the Hazard(s) and Hazardous Situation(s) associated with the Complaint Issue are documented in the Hazard Analysis (INX#100281795, Version # (9.0)). All complaint investigations are trended. There is no current trend alert documented.

Causality for "the button to apply the genotropin device is hard" was determined associated to device constituent of somatropin (malfunction).

The information on the batch/lot number for somatropin will be requested and submitted if and when received.