

SUSPECT ADVERSE REACTION REPORT

I. REACTION INFORMATION

1. PATIENT INITIALS (first, last) PRIVACY	1a. COUNTRY PANAMA	2. DATE OF BIRTH			2a. AGE 6 Years	3. SEX Female	3a. WEIGHT Unk	4-6 REACTION ONSET			8-12 CHECK ALL APPROPRIATE TO ADVERSE REACTION <input type="checkbox"/> PATIENT DIED <input type="checkbox"/> INVOLVED OR PROLONGED INPATIENT HOSPITALISATION <input type="checkbox"/> INVOLVED PERSISTENT OR SIGNIFICANT DISABILITY OR INCAPACITY <input type="checkbox"/> LIFE THREATENING
		Day	Month	Year				Day	Month	Year	
		PRIVACY						Unk			

7 + 13 DESCRIBE REACTION(S) (including relevant tests/lab data)
Event Verbatim [LOWER LEVEL TERM] (Related symptoms if any separated by commas)
painful injection [Injection site pain]
the medication could not be administered [Drug dose omission by device]
injection process was difficult because the pen was not friendly [Device difficult to use]
the needles of the pen were damaged [Needle issue]
the pen turned with difficulty/dosing button was jammed, it did not turned/it doesn't screw it on properly [Device mechanical jam]
80% of the medication was spilled. [Device leakage]
the numbers were not visible [Device image display issue]

(Continued on Additional Information Page)

II. SUSPECT DRUG(S) INFORMATION

14. SUSPECT DRUG(S) (include generic name) #1) Genotropin Pen (SOMATROPIN) Solution for injection #2) Genotropin Pen (SOMATROPIN (DEVICE CONSTITUENT)) Solution for injection {Lot # A143}		20. DID REACTION ABATE AFTER STOPPING DRUG? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
15. DAILY DOSE(S) #1) 0.8 mg, daily #2)	16. ROUTE(S) OF ADMINISTRATION #1) Unknown #2) Unknown	
17. INDICATION(S) FOR USE #1) Unknown #2) Unknown		21. DID REACTION REAPPEAR AFTER REINTRODUCTION? <input type="checkbox"/> YES <input type="checkbox"/> NO <input checked="" type="checkbox"/> NA
18. THERAPY DATES(from/to) #1) Unknown #2) Unknown	19. THERAPY DURATION #1) Unknown #2) Unknown	

III. CONCOMITANT DRUG(S) AND HISTORY

22. CONCOMITANT DRUG(S) AND DATES OF ADMINISTRATION (exclude those used to treat reaction)		
23. OTHER RELEVANT HISTORY. (e.g. diagnostics, allergies, pregnancy with last month of period, etc.) From/To Dates Type of History / Notes Description Unknown		

IV. MANUFACTURER INFORMATION

24a. NAME AND ADDRESS OF MANUFACTURER Pfizer S.A. Laura Arce Mora Avenida Escazú, Torre Lexus, piso 7. Escazú San jose, COSTA RICA		26. REMARKS
	24b. MFR CONTROL NO. PV202500046317	
24c. DATE RECEIVED BY MANUFACTURER 15-APR-2025	24d. REPORT SOURCE <input type="checkbox"/> STUDY <input type="checkbox"/> LITERATURE <input type="checkbox"/> HEALTH PROFESSIONAL <input checked="" type="checkbox"/> OTHER: Spontaneous	
DATE OF THIS REPORT 21-APR-2025	25a. REPORT TYPE <input checked="" type="checkbox"/> INITIAL <input type="checkbox"/> FOLLOWUP:	
		25b. NAME AND ADDRESS OF REPORTER NAME AND ADDRESS WITHHELD. NAME AND ADDRESS WITHHELD. NAME AND ADDRESS WITHHELD.

ADDITIONAL INFORMATION**7+13. DESCRIBE REACTION(S) continued**

Case Description: This is a spontaneous report received from a Consumer or other non HCP and a Nurse, Program ID: 164974.

A 6-year-old female patient received somatropin (GENOTROPIN PEN), (Batch/Lot number: unknown) at 0.8 mg daily, Device Lot Number: A143, Device Expiration Date: May2025. The patient's relevant medical history and concomitant medications were not reported.

The following information was reported: INJECTION SITE PAIN (non-serious), outcome "unknown", described as "painful injection"; DRUG DOSE OMISSION BY DEVICE (non-serious), outcome "unknown", described as "the medication could not be administered"; DEVICE DIFFICULT TO USE (non-serious), outcome "unknown", described as "injection process was difficult because the pen was not friendly"; NEEDLE ISSUE (non-serious), outcome "unknown", described as "the needles of the pen were damaged"; DEVICE MECHANICAL ISSUE (non-serious), outcome "unknown", described as "the pen turned with difficulty/dosing button was jammed, it did not turned/it doesn't screw it on properly"; DEVICE LEAKAGE (non-serious), outcome "unknown", described as "80% of the medication was spilled."; DEVICE INFORMATION OUTPUT ISSUE (non-serious), outcome "unknown", described as "the numbers were not visible". The action taken for somatropin was unknown.

Causality for "painful injection", "the medication could not be administered", "injection process was difficult because the pen was not friendly", "the needles of the pen were damaged", "the pen turned with difficulty/dosing button was jammed, it did not turned/it doesn't screw it on properly", "80% of the medication was spilled." and "the numbers were not visible" was determined associated to device constituent of somatropin (malfunction).

Additional information: Reporter stated the needles of the pen were damaged and the medication could not be administered because the pen turned with difficulty and the numbers were not visible, so it was not possible to set the dose. Reporter mentioned that she had two daughters and the injection process was difficult because the pen was not friendly and this led to a painful injection and she did not know if it was because of the pen or a bad administration technique. Reporter also stated that she set the medication and it did not screw correctly so 80% of the medication was spilled. Nurse stated the dosing button was jammed, it did not turned.